

Sampling and Analysis Summary Information for PAMS VOC Target Species

See Methods for VOCs on Next Page

Number of VOC Compounds = 60

Compound Name	IUPAC Name (if different)	Group Designation (note 1)	AIRS Parameter Number (note 1)	Boiling Point (degrees C) (note 2)	Volatility	CAS Number (note 2)	Sampling Method Alternatives (note 3)	Separator (note 4)	Detector (note 5)	EPA Ref. Desig. For Current Method (note 6)	Detection Limit (ppbv) (note 6)	Alternative Methods (possibly lower cost) (note 6)
1 Ethane		paraffin	43202	-88.5	Very vol.	74-84-0	Multi-adsorbent	GC	MS	TO-17	0.2-25	No alternative
2 Propane		paraffin	43204	-42	Very vol.	74-98-6	Multi-adsorbent	GC	MS	TO-17	0.2-25	No alternative
3 Isobutane	2-Methylpropane	paraffin	43214	-12	Very vol.	75-28-5	Can+ads or Can	GC	MS/FID	TO-15	0.2-25	TO-14A
4 n-Butane		paraffin	43212	0	Very vol.	106-97-8	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
5 Isopentane	2-Methylbutane	paraffin	43221	28	Very vol.	78-78-4	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
6 n-Pentane		paraffin	43220	36	Very vol.	109-66-0	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
7 Cyclopentane		paraffin	43242	49	Very vol.	287-92-3	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
8 2,2-Dimethylbutane		paraffin	43244	50	Med. vol.	75-83-2	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
9 2,3-Dimethylbutane		paraffin	43284	58	Med. vol.	79-29-8	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
10 2-Methylpentane		paraffin	43285	60	Med. vol.	107-83-5	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
11 3-Methylpentane		paraffin	43230	63	Med. vol.	96-14-0	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
12 n-Hexane		paraffin	43231	69	Med. vol.	110-54-3	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
13 Methylcyclopentane		paraffin	43282	72	Med. vol.	96-37-7	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
14 Cyclohexane		paraffin	43248	81	Med. vol.	110-82-7	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
15 2,4-Dimethylpentane		paraffin	43247	81	Med. vol.	108-08-7	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
16 2-Methyl hexane		paraffin	43263	90	Med. vol.	591-76-4	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
17 2,3-Dimethylpentane		paraffin	43291	90	Med. vol.	565-59-3	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
18 3-Methylhexane		paraffin	43249	92	Med. vol.	6131-24-4	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
19 2,2,4-Trimethylpentane		paraffin	43250	99	Med. vol.	540-84-1	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
20 n-Heptane		paraffin	43232	99	Med. vol.	142-82-5	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
21 Methylcyclohexane		paraffin	43261	101	Med. vol.	108-87-2	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
22 2,3,4-Trimethylpentane		paraffin	43252	114	Med. vol.	565-75-3	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
23 2-Methylheptane		paraffin	43960	118	Med. vol.	592-27-8	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
24 3-Methylheptane		paraffin	43253	119	Med. vol.	6131-25-5	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
25 n-Octane		paraffin	43233	126	Less vol.	111-65-9	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
26 n-Nonane		paraffin	43235	151	Less vol.	111-84-2	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
27 n-Decane		paraffin	43238	174	Less vol.	124-18-5	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
28 n-Undecane		paraffin	43954	196	Less vol.	1120-21-4	Can+ads/can	GC	MS/FID	TO-15	0.2-25	TO-14A
29 n-Dodecane		paraffin	43141	217	Less vol.	112-40-3	Can+ads	GC	MS	TO-15	0.2-25	No alternative
1 Acetylene	Ethyne	alkyne	43206	-85	Very vol.	74-86-2	Multi-adsorbent	GC	MS	TO-17	0.2-25	No alternative
1 Ethylene	Ethene	olefin	43203	-104	Very vol.	74-85-1	Multi-adsorbent	GC	MS	TO-17	0.2-25	No alternative
2 Propylene	1-Propene	olefin	43205	-48	Very vol.	115-07-1	Multi-adsorbent	GC	MS	TO-17	0.2-25	No alternative
3 1-Butene		olefin	43280	-6	Very vol.	106-98-9	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
4 trans-2-Butene		olefin	43216	1	Very vol.	624-64-6	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
5 cis-2-Butene		olefin	43217	4	Very vol.	590-18-1	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
6 1-Pentene		olefin	43224	30	Very vol.	109-67-1	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
7 Isoprene	2-Methyl-1,3-butadiene	olefin	43243	34	Very vol.	78-79-5	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
8 trans-2-Pentene		olefin	43226	36	Very vol.	646-04-8	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
9 cis-2-Pentene		olefin	43227	37	Very vol.	627-20-3	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
10 1-Hexene		olefin	43245	63	Med. vol.	592-41-6	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
1 Benzene		aromatic	45201	80	Med. vol.	71-43-2	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
2 Toluene	Methyl-benzene	aromatic	45202	111	Med. vol.	108-98-3	Can+ads/can/CMS	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-2
3 Ethylbenzene		aromatic	45203	136	Less vol.	100-41-4	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
4 meta-Xylene	1,3-Methyl-benzene	aromatic	45109	139	Less vol.	108-38-3	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
5 para-Xylene	1,4-Methyl-benzene	aromatic	45109	138	Less vol.	106-42-3	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
6 Styrene	Etheryl-benzene	aromatic	45220	145	Less vol.	100-42-5	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
7 ortho-Xylene	1,2-Methyl-benzene	aromatic	45204	145	Less vol.	95-47-6	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
8 Isopropylbenzene (cumene)	1-Methyl-ethyl-benzene	aromatic	45210	152	Less vol.	98-82-8	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
9 n-propylbenzene	Propyl-benzene	aromatic	45209	159	Less vol.	103-65-1	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
10 m-Ethyltoluene	1-Ethyl-3-methyl-benzene	aromatic	45212	161	Less vol.	620-14-4	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
11 p-Ethyltoluene	1-Ethyl-4-methyl-benzene	aromatic	45213	162	Less vol.	622-96-8	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
12 1,3,5-Trimethylbenzene		aromatic	45207	165	Less vol.	108-67-8	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
13 o-Ethyltoluene	1-Ethyl-2-methyl-benzene	aromatic	45211	165	Less vol.	611-14-3	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
14 1,2,4-Trimethylbenzene		aromatic	45208	169	Less vol.	95-63-6	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
15 1,2,3-Trimethylbenzene		aromatic	45225	176	Less vol.	526-73-8	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
16 m-Diethylbenzene	1,3-Diethyl-benzene	aromatic	45218	181	Less vol.	141-93-5	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
17 p-Diethylbenzene	1,2-Diethyl-benzene	aromatic	45219	184	Less vol.	105-05-5	Can+ads/can/cryog.	GC	MS/FID	TO-15	0.2-25	TO-14A/TO-3
1 Acetaldehyde	Ethanal	oxidized alcohols	43503	20	Very vol.	75-07-0	Cartridge/Liquid Impinger	HPLC	UV	TO-11A	0.5-100	TO-5
2 Acetone	2-Propanone	oxidized alcohols	43551	56	Med. vol.	67-64-1	Cartridge/Liquid Impinger	HPLC	UV	TO-11A	0.5-100	TO-5
3 Formaldehyde	Methanal	oxidized alcohols	43502	-20	Very vol.	50-00-0	Cartridge/Liquid Impinger	HPLC	UV	TO-11A	0.5-100	TO-5

note 1: See lists and discussion in "Technical Assistance Document for Sampling and Analysis of Ozone Precursors" EPA/600/R-98/161 (USEPA, Human Exposure and Atmospheric Sciences Division, Research Triangle Park, North Carolina, September, 1998, Section 2, pp. 5-7.

note 2: Boiling Points and CAS numbers are found in "CRC Handbook of Chemistry and Physics," 79th Edition, D. R. Lide, ed., Boca Raton, January, 1998, Section 3, pp. 3-1 ff.

note 3: At a simple level, sampling procedures fall into either canister techniques or adsorbent techniques. But the five methods, TO2,3,14A,15 and 17, provide for alternatives within these two categories. In addition, adsorbents vary with respect to breakthrough limits and VOC volatilities. The abbreviations shown include: Can = canister of any type, CMS = carbon molecular sieve adsorbent, Cry = cryogenic concentration technique (types vary), Ads = adsorbent of type other than CMS, including multisorbent tubes. Generally, it is assumed that most canister sampling methods are more costly than most adsorbent methods. However, complex multi-adsorbent cartridges can be costly.

note 4: Gas chromatograph is the designated separation method for both mass spectrometer and flame ionization methods.

note 5: Although mass spectrometer is the method of detection given for the most recent EPA methods, flame ionization is shown as an alternative detector for Methods TO-14A and TO-2.

note 6: Detailed descriptions of methods TO-1 through TO-17 are shown at <http://www.epa.gov/ttn/amt/c/airtox.html>.

METHODS FOR PAMS VOCs (note 1)

See VOC Species Information on Previous Page

Method Designation	Collector	Analyzer	Detector	Volatility category that method best matches (note 2)	Boiling pt. range (C) (note 1)	Most Appropriate Compounds	Detection limit (ppbv)	Cost comments and ratings: 1 = least costly (note 3)	Procedural Steps in Methods				
									1. Sample Collection	2. Sample Treatment	3. Sample Transfer	4. Separation	5. Detection, Identification, and Measurement
TO-1	Tenax cartridge	GC	MS	Less volatile	80 to 200	aromatic hydrocarbons, benzene, toluene, and xylene	0.01 to 100	2: MS is costly, but no canister required	1. Collect sample by drawing ambient air through Tenax cartridge.	2. Return to lab. Heat cartridge and purge with inert gas.	3. Transfer VOCs to cryog. trap, then heat trap for insertion of VOCs into GC.	4. Hold GC column at low temperature, then heat as VOCs are introduced.	5. Separate by GC and identify and measure by MS. ECD and FID are mentioned, but not identified as part of this method.
TO-2	Carbon molecular sieve cartridge	GC	MS FID	Medium volatile	-15 to 120	benzene, toluene	0.1 to 200	1: FID not as costly as MS, and no canister req.	1. Collect sample by drawing ambient air through CMS cartridge.	2. Return to lab and purge water vapor from cartridge with dry air and heated helium.	3. Transfer VOCs to cryog. loop (trap), then heat trap for insertion of VOCs into GC.	4. Hold GC column at low temperature, then heat as VOCs are introduced.	5. Separate by GC and identify and measure by MS. FID is identified as a possibly preferable for this method.
TO-3	Cryogenic canister	GC	FID	Medium volatile	-10 to 200	many VOCs	0.1 to 200	2: Cryog. canister system raises cost, but FID cheaper than MS	1. Collect sample by drawing ambient air through cryog. trap (container), e.g., immersed in liquid argon.	2. May use Nafion or other dryer before air goes into cryog. container.	3. No intermediate transfer.	4. Cryog. cont. intake valve is switched to GC column injection, possibly on site. Cont. is heated to 150 deg C.	5. Identify and measure compounds by FID (provides det. limits of 1 to 5 ng for many compounds).
TO-5	DNPH liquid impinger	HPLC	UV	Very volatile	-20 to 56	aldehydes and ketones	1 to 50	2: Uses HPLC	1. Draw ambient air into midget impinger containing 10 ml DNPH reagent	2. Place solution in vial and return to lab. Remove isooctane layer, extract aq.	3. Evaporate organic layers and dissolve residue in methanol.	4. Inject into HPLC.	5. Determine derivatives using UV detector at 370 nm.
TO-11A	DNPH Cartridge	HPLC	UV	Very volatile	-20 to 56	aldehydes and ketones	0.5 to 100	2: Similar to TO-5, but use of cartridge might be more costly	1. Draw ambient air into DNPH coated cartridge. Place cartridge in glass vial and seal.	2. Return to lab. Remove cartridge and wash with acetonitrile.	3. No further processing needed.	4. Acetonitrile solution is diluted and injected into HPLC.	5. Determine derivative by UV detection at 350 nm.
TO-14A	Canister / cryog. trap	GC	FID/ECD or MS	Medium volatile (covers almost all VOCs)	-29 to 213	non-polar VOCs	0.2 to 25	2: Canister system req., FID optional	1. Draw ambient air into canister (e.g. 6L) equipped with flow control device.	2. Return to lab. Dry with Nafion dryer or alternative.	3. Transfer VOCs to cryog loop (trap), then heat trap for insertion of VOCs into GC.	4. Separation in GC for transfer either to MS or to combination-detector system.	5. TO-14A describes either a two-way MS system (SCAN versus SIM) or a three-way FID-PID).
TO-15	Canister / sorbent trap	GC	MS	Medium volatile (covers almost all VOCs)	-50 to 240	polar/non-polar VOCs	0.2 to 25	3: Canister plus solid adsorbent with MS	1. Draw ambient air into canister (e.g. 6L) equipped with flow control device.	2. Return to lab. Pass sample through multisorbent packed tube. Purge water vapor with helium.	3. Cryog. trap concentrator optional.	4. Separation in GC.	5. identify and measure compounds by MS
TO-16	none	FTIR, open path	Infra-red spectrom.	Less volatile (covers med. also)	25 to 500	polar/non-polar VOCs		2: No sampling system req., but complex field equip.	1. No specific sampling system. All of the air in the line of the FTIR is "sampled".	2. none	3. none	4. none	6. Identify and measure compounds in open air by FTIR.
TO-17	Adsorbent tube	GC	MS	Very volatile (covers med. also)	-60 to 200	polar/non-polar VOCs	0.2 to 25	3: Uses multisorbent cartridge and MS	1. Draw ambient air through a multisorbent packed tube.	2. Seal and pack tube. Return to lab. Tube may be stored before analysis.	3. Transfer VOCs to intermediate adsorbent trap or directly to GC, by heating sampling tube.	4. Separation in GC.	5. Identify and measure compounds by MS.

Note 1 Most of the information in this table is from the EPA * Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air, 2nd Edition, EPA/625/R-96/010b, January 1999, available at the AMTIC webpage <http://www.epa.gov/ttn/amtic/airtox.html>, or from descriptions of the individual Methods, available at the same webpage.

Note 2 Temperature ranges for the methods are found in the Compendium referenced above, Table 2, pages 5-10, or in the descriptions of the Methods, where the temperature range for the Method is inferred from tests for detection of VOCs Boiling Points and CAS numbers are found in "CRC Handbook of Chemistry and Physics," 76th Edition, D. R. Lide, ed., Boca Raton, January 1995, Section 3, pp.3-1ff.

Effective temperature ranges for adsorbents are found in the description for TO-17, Table 1, pp. 17-33 to 17-44

FOR VERY VOLATILE VOCs: (BP < 50) Choose an adsorbent (multisorbent) with capability of adsorbing in the required BP range. Then choose a TO Method with an adsorbent sampling procedure.

FOR MEDIUM VOLATILE VOCs: (50 < BP < 120) Choose either a canister or an adsorbent system which covers the BP range as precisely as possible, so as to avoid the cost of excess capability.

FOR LESS VOLATILE VOCs: (120 < BP) Choose either a canister or an adsorbent system of lowest possible cost.

Note 3 The assumptions underlying the cost comments are, that in general, canister sampling is more costly than adsorbent because of the equipment required for field air intake, and that MS is a more costly method than FID because of the higher equipment cost. However, there are always tradeoffs, for example, between equipment cost and personnel training costs. And some sorbent cartridges may well be as costly as the comparable canister equipment.

	HIGHLY VOLATILE	MEDIUM VOLATILE	LESS VOLATILE
SAMPLING: canister adsorbent cann/ads	TO-14A(?) TO-17 TO-15 (?)	TO-14A TO-2 TO-15	none TO-3 TO-1
DETECTION: mass spec flame ion.	TO-15, TO-17 TO-14A (?)	TO-2, TO-14A TO-2, TO-14A	TO-1 TO-3
	(?) = not optimal match of ranges.		